



QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

File
LR1409

CERTIFICATE OF COMPLIANCE
(ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Casella (Ideal Industries Ltd)
Address	Regent House, Wolseley Road, Kempston, Bedford MK42 7JY
Project Number	LR1409-3
Product	Personal Air Sampling Pump
Model Number	VAPex
Ratings	Class I, Zone 0, AEx ia IIC T4 Ga Ex ia IIC T4 Ga Ta=-20°C to +45°C
Applicable Standards	CSA C22.2 No. 60079-0-19, CSA C22.2 No. 60079-11-14, CSA C22.2 No. 61010-1, 3rd Edition, UL 60079-0:19, UL 60079-11:2013, UL 61010-1, 3rd Edition
Factory/Manufacturing Location	Same as above

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Dave Adams

Signature: 

Date: August 14, 2020