



QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

File LR1409

CERTIFICATE OF COMPLIANCE
(ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Casella Holdings Limited	
Address	Regent House, Wolseley Road Kempston, Bedford Bedfordshire MK42 7JY United Kingdom	
Project Number	LR1409-1R3	
Product	Personal Sound Exposure Meter (Dosimeter)	
Model Number	dBadge2	
Ratings	Docking Station (207107B) Charging Um: 60 V Each Docking Station Uo: 5.36 V Io: 213 mA USB Um: 60 V Uo: 1.0 V Io: 170 mA dBadge2 Ui: 6.51 V	
Markings	Canada Ex ia IIC T4 Ga Ex ia IIIC T135°C Da -20°C to +45°C	US Class I Zone 0, AEx ia IIC T4 Ga Zone 20, AEx ia IIIC T135°C Da -20°C to +45°C
Applicable Standards	CAN/CSA C22.2 No. 60079-0:19 CAN/CSA C22.2 No. 60079-11:14 CSA C22.2 No. 61010-1-12 ANSI/UL 60079-0:2019 ANSI/UL 60079-11:2013 UL 61010:2015, Ed 3	
Factory/Manufacturing Location	Casella Holdings Limited Regent House, Wolseley Road, Kempston, Bedford, Bedfordshire MK42 7JY UK	

Statement of Compliance: The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



Issued By: Rob Kohuch, P. Eng.

Signature: *Rob Kohuch*

Date: November 1, 2022